

Sampler Specification

Specification Number

675 8920H450

Revision

ion 01

Part No. 675 8920H450

Description SteriWare LiquiDip Mono – 450mm

Material of Construction: High Density Polyethylene (HDPE) - virgin

Conforms to FDA CFR 177.1520

• Conforms to EU Regulations 10/2011

• Conforms to EC Regulations 1935/2004





Method of Construction: Ultrasonic welding

Overall Length 430mm

Maximum Diameter: 26mm

Outer Diameter of Tube: 21mm

Diameter of hole in tip: 6mm

Nominal Weight of Sampler: 50g

Assembly & Packing Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE bag)

Method of Sterilisation: N/A

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature